



# 龙子湖新能源实验室 大型仪器设备技术验收报告

设备名称：400MHz 固体核磁共振波谱仪

使用部门：公共科学技术中心

验收负责人：李行

验收日期：2024.11.30

龙子湖新能源实验室综合事务部

## 设备基本情况登记


设备名称		400MHz 固体核磁共振波谱仪		
规格型号		AVANCE NEO 400	数量	1
生产厂商		布鲁克	供应商	布鲁克（北京）科技有限公司
出厂日期		2024年8月	设备经费来源	自筹
合同号		豫财单一采购-2023-134	到货日期	2024年8月16日
安装使用地点		龙子湖新能源实验室2号楼北楼123		
价格	人民币	6010000.00	外 币	
使用责任人		李行	联系电话	18437911559
设备随机资料				
序号	名 称		份数	备注
1	超导磁体		1	
2	机柜		1	
3	$^1\text{H}/(^{15}\text{N}-^{31}\text{P})$ 1.9mm 双共振固体探头		1	
4	$^1\text{H}/\text{X}/\text{Y}$ 4mm 三共振固体探头		1	
5	$^1\text{H}/(^{13}\text{C}-^{109}\text{Ag})$ 7mm 双共振固体探头		1	
6	PC 工作站		1	
7	NMR 软件		1	
8	标准样品 1 套		1	
9	超导磁体用液氮真空输液管		1	

## 设备安装调试记录

记录人	李行	安装调试时间	2024.8.27
<p>(安装、调试、运行过程及结果等记录、与厂商代表洽谈情况等)</p> <p>2024年8月16日, 由布鲁克(北京)科技有限公司运输到龙子湖新能源实验室2号楼北楼123房间, 经验收, 设备无损坏, 设备数量与发货清单一致。</p> <p>2024年8月27日, 由布鲁克厂家委派方勇工程师对设备进行安装、调试。</p> <ol style="list-style-type: none"><li>1. 2024年8月27日 安装磁体;</li><li>2. 2024年8月28日至9月1日 安装磁体部件;</li><li>3. 2024年9月2日 往磁体液氦口加液氮, 进行液氮预冷;</li><li>4. 2024年9月4日 往磁体液氦口加液氮;</li><li>5. 2024年9月5日 安装BST探头;</li><li>6. 2024年9月6日 安装工作站和信号放大器, 进行调谐;</li><li>7. 2024年9月9日 磁体进行升场;</li><li>8. 2024年9月10日 低温匀场;</li><li>9. 2024年9月18日 安装氮气分离器;</li><li>10. 2024年9月19日 调试1.9mm、4mm和7mm探头。</li><li>11. 2024年9月25日至27日 进行固体核磁培训。</li><li>12. 经过一段时间的使用测试, 设备运行正常。</li></ol> <p><b>结果:</b></p> <p>设备及附件均已到货, 设备已安装并进行基础培训。</p>			



数据结果显示:



# 设备验收记录

(仪器设备性能及技术指标与合同规定要求符合程度、验证方式方法、验证标准以及测试资料

(包括曲线、图纸、照片等) 并加以说明)

## Installation Protocol - NMR Spectrometer

### NMR Probes

Description	Probe ID	Inspection Lot	Status
PH MASDVT400W1 BL4 X/Y/H	H8906_0110	2024-09-19	pass
PH MASDVT400W1 BL4 X/Y/H	H8906_0110	2024-09-19_TripleMode_HCN	pass
PH MASDVT400W1 BL4 X/Y/H	H8906_0110	2024-09-19_TripleMode_HPC	pass
PH MASDVT400W1 BL1.9 N-P/H	H171392_0003	2024-09-20	pass
PH MAS DVT 400W1 BL7 109Ag-AI/F-H	H132625_0012	2024-09-19	pass

Copies of all spectra (default and additional) are included in customer's PDF report.

### Installation Checklist

Installation	pass	fail	n/a	Optional Components	pass	fail	n/a
All connections and grounding	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Sample Changer	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
All firmware	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	MAS controller	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Contab for required nuclei	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	High power equipment	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Lift / spin calibration	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	LC-NMR	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Software licenses	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Liquid Handler SamplePro Tube	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
He / N2 log files activated	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Micro-Imaging	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
MICS installed	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Diffusion	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Remote monitoring enabled	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	CryoProbe / CryoPlatform	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Customer agreed in activating	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	BNL / BSNL	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
AutoDiagnose	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Additional cooling/heating units (like BCU1 / BCU2)	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
AutoDiagnose successfully activated	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	LT-MAS (Low Temperature MAS equipment)	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
Customer agreed in installing	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Gyrottron magnet and DNP console	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
LabScape Pro Care	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
LabScape Pro Care successfully installed	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
<b>Customer Training</b>	<b>pass</b>	<b>fail</b>	<b>n/a</b>				
Basic safety	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Magnet safety and refilling	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Handling of cryogenic liquids	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Hardware overview	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Console on/off operation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Basic operation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Troubleshooting	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Backup (nmr_save, images)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Introduction to IconNMR	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
Assure-SST / Performance check	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
CryoProbe	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
Handling / cleaning of probe	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
He cylinder exchange	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
He compressor cooling	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
RF heating / power limits	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
RF routing	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>				
<b>Acceptance and Warranty</b>	<b>pass</b>	<b>fail</b>	<b>n/a</b>				
Explanation of warranty	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Spectrometer documentation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				
Customer support hotlines	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				

# 实验结果:

NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: HS905\_0110 PH MASDVT400W1 BL4 XY/H  
 Sample: Potassium Bromide (KBr, 80 ul) (Z151220)  
 Magic Angle setting, MAS (NPT\_79Br\_MAS\_magicAngle, spin rate 5000 Hz)

Line width main [achieved/rated]: [142 <= 180] <pass>  
 Line width of side band number 8 (@ -40007 Hz) [achieved/rated]: [133 <= 200] <pass>




Bruker BioSpin

## NPT\_79Br\_MAS\_magicAngle

```

===== Data Parameters
NAME: NPT_79Br_MAS_magicAngle
EXPNO: 1
PROCNO: 1
F2 - Acquisition Parameters
Date_ 20150725
Time: 10:55 h
PROBHD: Avance
PROBHD2: 4001_0110 PH
PULPROG: zgpg30
TD: 65536
SOLVENT: H2O
NS: 16
DS: 4
SWH: 47000.000 Hz
FIDRES: 0.001700 Hz
AQ: 0.0004600 sec
RG: 321
AQ2: 0.0004600 sec
RG2: 321
RG3: 321
RG4: 321
RG5: 321
RG6: 321
RG7: 321
RG8: 321
RG9: 321
RG10: 321
RG11: 321
RG12: 321
RG13: 321
RG14: 321
RG15: 321
RG16: 321
RG17: 321
RG18: 321
RG19: 321
RG20: 321
RG21: 321
RG22: 321
RG23: 321
RG24: 321
RG25: 321
RG26: 321
RG27: 321
RG28: 321
RG29: 321
RG30: 321
RG31: 321
RG32: 321
RG33: 321
RG34: 321
RG35: 321
RG36: 321
RG37: 321
RG38: 321
RG39: 321
RG40: 321
RG41: 321
RG42: 321
RG43: 321
RG44: 321
RG45: 321
RG46: 321
RG47: 321
RG48: 321
RG49: 321
RG50: 321
RG51: 321
RG52: 321
RG53: 321
RG54: 321
RG55: 321
RG56: 321
RG57: 321
RG58: 321
RG59: 321
RG60: 321
RG61: 321
RG62: 321
RG63: 321
RG64: 321
RG65: 321
RG66: 321
RG67: 321
RG68: 321
RG69: 321
RG70: 321
RG71: 321
RG72: 321
RG73: 321
RG74: 321
RG75: 321
RG76: 321
RG77: 321
RG78: 321
RG79: 321
RG80: 321
RG81: 321
RG82: 321
RG83: 321
RG84: 321
RG85: 321
RG86: 321
RG87: 321
RG88: 321
RG89: 321
RG90: 321
RG91: 321
RG92: 321
RG93: 321
RG94: 321
RG95: 321
RG96: 321
RG97: 321
RG98: 321
RG99: 321
RG100: 321
===== Processing parameters
SI: 32768
SF: 100.626100 MHz
WDW: EM
SSB: 0
LB: 0 Hz
GB: 0 Hz
PC: 1
RG: 321
AQ: 0.0004600 sec
RG2: 321
RG3: 321
RG4: 321
RG5: 321
RG6: 321
RG7: 321
RG8: 321
RG9: 321
RG10: 321
RG11: 321
RG12: 321
RG13: 321
RG14: 321
RG15: 321
RG16: 321
RG17: 321
RG18: 321
RG19: 321
RG20: 321
RG21: 321
RG22: 321
RG23: 321
RG24: 321
RG25: 321
RG26: 321
RG27: 321
RG28: 321
RG29: 321
RG30: 321
RG31: 321
RG32: 321
RG33: 321
RG34: 321
RG35: 321
RG36: 321
RG37: 321
RG38: 321
RG39: 321
RG40: 321
RG41: 321
RG42: 321
RG43: 321
RG44: 321
RG45: 321
RG46: 321
RG47: 321
RG48: 321
RG49: 321
RG50: 321
RG51: 321
RG52: 321
RG53: 321
RG54: 321
RG55: 321
RG56: 321
RG57: 321
RG58: 321
RG59: 321
RG60: 321
RG61: 321
RG62: 321
RG63: 321
RG64: 321
RG65: 321
RG66: 321
RG67: 321
RG68: 321
RG69: 321
RG70: 321
RG71: 321
RG72: 321
RG73: 321
RG74: 321
RG75: 321
RG76: 321
RG77: 321
RG78: 321
RG79: 321
RG80: 321
RG81: 321
RG82: 321
RG83: 321
RG84: 321
RG85: 321
RG86: 321
RG87: 321
RG88: 321
RG89: 321
RG90: 321
RG91: 321
RG92: 321
RG93: 321
RG94: 321
RG95: 321
RG96: 321
RG97: 321
RG98: 321
RG99: 321
RG100: 321
    
```



NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: HS905\_0110 PH MASDVT400W1 BL4 XY/H  
 Sample: Potassium Bromide (KBr, 80 ul) (Z151220)  
 Maximum spin rate testing, MAS (NPT\_79Br\_MAS\_maxSpinRate, spin rate 15000 Hz)  
 Determination of spinning stability for 180 s  
 Pressure values in mbar: DrivePressure=1431/BearingPressure=3176/BearingSensePressure=3196/SupplyPressure=5886/SystemPressure=5325

Spin rate at maximum deviation [measured]: @ MASR 15000 Hz [15002 Hz]  
 Maximum deviation [achieved/rated]: @ MASR 15000 Hz [2 Hz <= 15 Hz] <pass>




Bruker BioSpin

## NPT\_79Br\_MAS\_maxSpinRate

```

===== Data Parameters
NAME: NPT_79Br_MAS_maxSpinRate
EXPNO: 1
PROCNO: 1
F2 - Acquisition Parameters
Date_ 20150725
Time: 10:55 h
PROBHD: Avance
PROBHD2: 4001_0110 PH
PULPROG: zgpg30
TD: 65536
SOLVENT: H2O
NS: 16
DS: 4
SWH: 47000.000 Hz
FIDRES: 0.001700 Hz
AQ: 0.0004600 sec
RG: 321
AQ2: 0.0004600 sec
RG2: 321
RG3: 321
RG4: 321
RG5: 321
RG6: 321
RG7: 321
RG8: 321
RG9: 321
RG10: 321
RG11: 321
RG12: 321
RG13: 321
RG14: 321
RG15: 321
RG16: 321
RG17: 321
RG18: 321
RG19: 321
RG20: 321
RG21: 321
RG22: 321
RG23: 321
RG24: 321
RG25: 321
RG26: 321
RG27: 321
RG28: 321
RG29: 321
RG30: 321
RG31: 321
RG32: 321
RG33: 321
RG34: 321
RG35: 321
RG36: 321
RG37: 321
RG38: 321
RG39: 321
RG40: 321
RG41: 321
RG42: 321
RG43: 321
RG44: 321
RG45: 321
RG46: 321
RG47: 321
RG48: 321
RG49: 321
RG50: 321
RG51: 321
RG52: 321
RG53: 321
RG54: 321
RG55: 321
RG56: 321
RG57: 321
RG58: 321
RG59: 321
RG60: 321
RG61: 321
RG62: 321
RG63: 321
RG64: 321
RG65: 321
RG66: 321
RG67: 321
RG68: 321
RG69: 321
RG70: 321
RG71: 321
RG72: 321
RG73: 321
RG74: 321
RG75: 321
RG76: 321
RG77: 321
RG78: 321
RG79: 321
RG80: 321
RG81: 321
RG82: 321
RG83: 321
RG84: 321
RG85: 321
RG86: 321
RG87: 321
RG88: 321
RG89: 321
RG90: 321
RG91: 321
RG92: 321
RG93: 321
RG94: 321
RG95: 321
RG96: 321
RG97: 321
RG98: 321
RG99: 321
RG100: 321
===== Processing parameters
SI: 32768
SF: 100.626100 MHz
WDW: EM
SSB: 0
LB: 0 Hz
GB: 0 Hz
PC: 1
RG: 321
AQ: 0.0004600 sec
RG2: 321
RG3: 321
RG4: 321
RG5: 321
RG6: 321
RG7: 321
RG8: 321
RG9: 321
RG10: 321
RG11: 321
RG12: 321
RG13: 321
RG14: 321
RG15: 321
RG16: 321
RG17: 321
RG18: 321
RG19: 321
RG20: 321
RG21: 321
RG22: 321
RG23: 321
RG24: 321
RG25: 321
RG26: 321
RG27: 321
RG28: 321
RG29: 321
RG30: 321
RG31: 321
RG32: 321
RG33: 321
RG34: 321
RG35: 321
RG36: 321
RG37: 321
RG38: 321
RG39: 321
RG40: 321
RG41: 321
RG42: 321
RG43: 321
RG44: 321
RG45: 321
RG46: 321
RG47: 321
RG48: 321
RG49: 321
RG50: 321
RG51: 321
RG52: 321
RG53: 321
RG54: 321
RG55: 321
RG56: 321
RG57: 321
RG58: 321
RG59: 321
RG60: 321
RG61: 321
RG62: 321
RG63: 321
RG64: 321
RG65: 321
RG66: 321
RG67: 321
RG68: 321
RG69: 321
RG70: 321
RG71: 321
RG72: 321
RG73: 321
RG74: 321
RG75: 321
RG76: 321
RG77: 321
RG78: 321
RG79: 321
RG80: 321
RG81: 321
RG82: 321
RG83: 321
RG84: 321
RG85: 321
RG86: 321
RG87: 321
RG88: 321
RG89: 321
RG90: 321
RG91: 321
RG92: 321
RG93: 321
RG94: 321
RG95: 321
RG96: 321
RG97: 321
RG98: 321
RG99: 321
RG100: 321
    
```




NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: H9906\_0110 PH MASDVT400W1 BL4 X/Y/H  
 Sample: Adamantane (50 ug) (Z151221)  
 P90 1H pulse calibration, MAS (NPT\_1H\_MAS\_p90dat\_1h, spin rate 15000 Hz)  
 ATTENTION: Updated PROSCL Tables with [2.50 us @ 87.0 W]  
 P90 MAS 1H pulse [achieved/rated]: @ 90.1 W [2.46 us @ 2.50 us] <pass>



Bruker BioSpin

NPT\_1H\_MAS\_p90dat\_1h



```
===== Data Parameters =====
NAME: NPT_1H_MAS_p90dat_1h
EXPNO: 1
PROCNO: 1
F2 - Acquisition Parameters
Date_ 20240919
Time: 11:27:05
PULPROG: zgpg30
PROBHD: zgpg30_1h
PCPDPRG2: zgpg30
PCPDPRG1: zgpg30
SOLVENT: none
AQ: 0.00000000
RG: 0
SI: 0
SF: 15000.00000000
FIDRES: 0.00000000
AQRES: 0.00000000
SFO: 400.1800000000000
NUC1: 13C
NUC2: 1H
P1: 0.00000000
P2: 0.00000000
P3: 0.00000000
P4: 0.00000000
P5: 0.00000000
P6: 0.00000000
P7: 0.00000000
P8: 0.00000000
P9: 0.00000000
P10: 0.00000000
P11: 0.00000000
P12: 0.00000000
P13: 0.00000000
P14: 0.00000000
P15: 0.00000000
P16: 0.00000000
P17: 0.00000000
P18: 0.00000000
P19: 0.00000000
P20: 0.00000000
===== Processing parameters =====
SI: 0
SF: 400.1800000000000
WDW: EM
SSB: 0
GB: 0
PC: 0
SC: 0
MC: 0
MS: 0
AS: 0
RN: 0
DC: 0
DS: 0
HS: 0
HT: 0
AQ: 0.00000000
RG: 0
SI: 0
SF: 15000.00000000
FIDRES: 0.00000000
AQRES: 0.00000000
SFO: 400.1800000000000
NUC1: 13C
NUC2: 1H
P1: 0.00000000
P2: 0.00000000
P3: 0.00000000
P4: 0.00000000
P5: 0.00000000
P6: 0.00000000
P7: 0.00000000
P8: 0.00000000
P9: 0.00000000
P10: 0.00000000
P11: 0.00000000
P12: 0.00000000
P13: 0.00000000
P14: 0.00000000
P15: 0.00000000
P16: 0.00000000
P17: 0.00000000
P18: 0.00000000
P19: 0.00000000
P20: 0.00000000
```

```
===== 13C Data Parameters =====
NAME: NPT_13C_MAS_sino_13c
EXPNO: 1
PROCNO: 1
F2 - Acquisition Parameters
Date_ 20240919
Time: 11:27:05
PULPROG: zgpg30
PROBHD: zgpg30_1h
PCPDPRG2: zgpg30
PCPDPRG1: zgpg30
SOLVENT: none
AQ: 0.00000000
RG: 0
SI: 0
SF: 15000.00000000
FIDRES: 0.00000000
AQRES: 0.00000000
SFO: 400.1800000000000
NUC1: 13C
NUC2: 1H
P1: 0.00000000
P2: 0.00000000
P3: 0.00000000
P4: 0.00000000
P5: 0.00000000
P6: 0.00000000
P7: 0.00000000
P8: 0.00000000
P9: 0.00000000
P10: 0.00000000
P11: 0.00000000
P12: 0.00000000
P13: 0.00000000
P14: 0.00000000
P15: 0.00000000
P16: 0.00000000
P17: 0.00000000
P18: 0.00000000
P19: 0.00000000
P20: 0.00000000
```

NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: H9906\_0110 PH MASDVT400W1 BL4 X/Y/H  
 Sample: Adamantane (50 ug) (Z151221)  
 13C sensitivity, MAS (NPT\_13C\_MAS\_sino\_13c, spin rate 15000 Hz)



Bruker BioSpin

NPT\_13C\_MAS\_sino\_13c

```
# Thu Sep 19 03:38:08 2024
##PROBHD=H9906_0110
##PROBNAME=PH MASDVT400W1 BL4 X/Y/H
##SOLVENT=none
##CRYSTALID=D315-54-6922
##SHIMID=312532
#
# Active Shim Gradients
#
Z -17650
Z2 0
Z3 0
Z4 0
Z5 0
X 0
X2 0
XZ2 0
Y 0
YZ 0
YZ2 0
XY 0
XYZ 0
(XZ-YZ) -66000
(XZ-YZ)2 0
X3 0
Y3 0
#
IEEE64_VERSION_CODE 1
#
# Shim currents
#
SHIM SETTING [ 1] -8560.25000000
SHIM SETTING [ 2] 0.00000000
SHIM SETTING [ 3] 0.00000000
SHIM SETTING [ 4] 0.00000000
SHIM SETTING [ 5] 0.00000000
SHIM SETTING [ 6] 0.00000000
SHIM SETTING [ 7] 0.00000000
SHIM SETTING [ 8] 0.00000000
SHIM SETTING [ 9] -34716.00000000
SHIM SETTING [10] 0.00000000
SHIM SETTING [11] 0.00000000
SHIM SETTING [12] 0.00000000
SHIM SETTING [13] 0.00000000
SHIM SETTING [14] 2741.90000052
SHIM SETTING [15] 13394.48461080
SHIM SETTING [16] 0.00000000
SHIM SETTING [17] 1323.00000256
SHIM SETTING [18] 0.00000000
SHIM SETTING [19] 0.00000000
SHIM SETTING [20] 0.00000000
```

~/NPT\_shimfile\_right.txt

```
===== Data Parameters =====
NAME: NPT_13C_MAS_sino_13c
EXPNO: 1
PROCNO: 1
F2 - Acquisition Parameters
Date_ 20240919
Time: 11:27:05
PULPROG: zgpg30
PROBHD: zgpg30_1h
PCPDPRG2: zgpg30
PCPDPRG1: zgpg30
SOLVENT: none
AQ: 0.00000000
RG: 0
SI: 0
SF: 15000.00000000
FIDRES: 0.00000000
AQRES: 0.00000000
SFO: 400.1800000000000
NUC1: 13C
NUC2: 1H
P1: 0.00000000
P2: 0.00000000
P3: 0.00000000
P4: 0.00000000
P5: 0.00000000
P6: 0.00000000
P7: 0.00000000
P8: 0.00000000
P9: 0.00000000
P10: 0.00000000
P11: 0.00000000
P12: 0.00000000
P13: 0.00000000
P14: 0.00000000
P15: 0.00000000
P16: 0.00000000
P17: 0.00000000
P18: 0.00000000
P19: 0.00000000
P20: 0.00000000
```





NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: H132625\_0012 PH MAS DVT 400W1 BL7 109Ag-AIF-H  
 Sample: Potassium Bromide (KBr, 234 ul) (Z151210)  
 Magic Angle setting, MAS (NPT\_79Br\_MAS\_magicAngle, spin rate 5000 Hz)

Line width main [achieved/rated]: [127 <= 180] <pass>  
 Line width of side band number 8 (@ -40005 Hz) [achieved/rated]: [124 <= 200] <pass>




Bruker BioSpin

NPT\_79Br\_MAS\_magicAngle

```

===== Data Parameters
NAME: NPT_79Br_MAS_magicAngle
EXPNO: 1
PROCNO: 1
PROCRES: 1
AQ: 0.00000000
RG: 327.50000000
SF: 400.18000000
WDW: EM
SSB: 0
LB: 0.00000000
GB: 0
PC: 0.00000000
SC: 0
DC: 0
AS: 0
SI: 0
SF2: 0.00000000
WDW2: EM
SSB2: 0
LB2: 0.00000000
GB2: 0
PC2: 0.00000000
SC2: 0
DC2: 0
AS2: 0
SI2: 0
===== Processing parameters
SI: 0
SF: 400.18000000
WDW: EM
SSB: 0
LB: 0.00000000
GB: 0
PC: 0.00000000
SC: 0
DC: 0
AS: 0
SI: 0
  
```



NMR TEST ACCEPTANCE \*\*\* System: AV NEO (400.18 MHz) \*\*\* TopSpin 4.4.0  
 Probe: H132625\_0012 PH MAS DVT 400W1 BL7 109Ag-AIF-H  
 Sample: Potassium Bromide (KBr, 234 ul) (Z151210)  
 Maximum spin rate testing, MAS (NPT\_79Br\_MAS\_maxSpinRate, spin rate 7000 Hz)  
 Determination of spinning stability for 180 s  
 Pressure values in mbar: DrivePressure=1400/BearingPressure=3164/BearingSensePressure=3187/SupplyPressure=5825/SystemPressure=5269

Spin rate at maximum deviation [measured]: @ MASR 7000 Hz [6998 Hz]  
 Maximum deviation [achieved/rated]: @ MASR 7000 Hz [2 Hz <= 7 Hz] <pass>




Bruker BioSpin

NPT\_79Br\_MAS\_maxSpinRate

```

===== Data Parameters
NAME: NPT_79Br_MAS_maxSpinRate
EXPNO: 1
PROCNO: 1
PROCRES: 1
AQ: 0.00000000
RG: 327.50000000
SF: 400.18000000
WDW: EM
SSB: 0
LB: 0.00000000
GB: 0
PC: 0.00000000
SC: 0
DC: 0
AS: 0
SI: 0
SF2: 0.00000000
WDW2: EM
SSB2: 0
LB2: 0.00000000
GB2: 0
PC2: 0.00000000
SC2: 0
DC2: 0
AS2: 0
SI2: 0
===== Processing parameters
SI: 0
SF: 400.18000000
WDW: EM
SSB: 0
LB: 0.00000000
GB: 0
PC: 0.00000000
SC: 0
DC: 0
AS: 0
SI: 0
  
```





## 设备验收小组意见

验收小组 结论意见	同意验收	
	验收小组组长（签字） <span style="font-size: 24px;">宋玉婷</span>	年 月 日
设备验收小组成员：		
姓名	单位	签字
宋玉婷	龙子湖新能源实验室	<span style="font-size: 24px;">宋玉婷</span>
吕婉雪	龙子湖新能源实验室	<span style="font-size: 24px;">吕婉雪</span>
李申	龙子湖新能源实验室	<span style="font-size: 24px;">李申</span>
郑双双	龙子湖新能源实验室	<span style="font-size: 24px;">郑双双</span>
李行	龙子湖新能源实验室	<span style="font-size: 24px;">李行</span>

